

## Design and development of a $\mu$ PAD device for magnesium determination in saliva

**Juliana I. S. Aguiar, Raquel B. R. Mesquita\*, António O. S. S. Rangel**

Universidade Católica Portuguesa, CBQF – Centro de Biotecnologia e Química Fina – Laboratório

Associado, Escola Superior de Biotecnologia, Rua Arquiteto Lobão Vital, 172, 4200-374 Porto, Portugal

\*rmesquita@porto.ucp.pt

Magnesium is the second most important intracellular cation and the fourth most prevalent. Disorders involving magnesium are categorized into two groups: hypomagnesemia (magnesium deficiency) and hypermagnesemia. Therefore, the determination of magnesium can be of great interest, because it helps in the clinical context and epidemiological research.

In this context, the objective of propose work was to develop a microfluidic paper-based analytical device ( $\mu$ PAD) for magnesium quantification in saliva. There are two main advantages on using saliva as an alternative biological sample namely being easy to collect and non-invasive. The  $\mu$ PADs have several characteristics which make them ideally suited to conduct on-location analyte determinations: there is no need for specialized skills or expensive equipment; all materials are cheap and easily obtained; the overall size is ideally for on-hand determination; and they can be discarded after use in a disposable approach. Another advantage of these devices is that they use small amounts of sample and reagent, still providing both immediate and accurate results.

The developed  $\mu$ PAD was based on the color reaction between magnesium and eriochrome cyanine dye. The  $\mu$ PAD was assembled with the reagent, and then the sample was added, followed by the sealing of the device. After 5 min (for the reaction to occur) a scanner was used to obtain a high resolution image of the detection zone and then the colour intensity was measured (Image J software) to established calibration curves.




Fig.1. Schematic representation of the developed  $\mu$ PAD (a) and photo of the two  $\mu$ PADs representing a calibration curve (b)

ACKNOWLEDGMENTS: R.B.R. Mesquita thanks FCT - Fundação para a Ciência e a Tecnologia for the grant SFRH/BDP/112032/2015. This work was supported by National Funds from FCT through project UID/Multi/50016/2013.